

APPLICATIONS

Corticosteroid Extraction From Plasma Using Strata[®] DE Supported Liquid Extraction (SLE)

Jessica Detsch, Zeshan Aqeel and Matt Brusius
Phenomenex, Inc., 411 Madrid Ave., Torrance, CA 90501 USA

Overview

Sensitive analytical tools are necessary to detect and confirm the presence of corticosteroids. In this application, we develop a Supported Liquid Extraction (SLE) method using Strata DE (diatomaceous earth) for a wide range of corticosteroid compounds from plasma, which are then analyzed by LC-MS/MS.

All compounds in the suite provided recovery greater than 90 %, with the exception of Triamcinolone. Triamcinolone is the most polar compound in the suite and is simply too hydrophilic to be extracted by DCM. Acceptable recoveries can be obtained by changing to ethyl acetate as an elution solvent. All compounds show a % CV of less than 12 %.

Pre-treatment

Dilute 100 µL of spiked plasma (125 ng/mL) with 200 µL of Water

SLE Protocol

96-Well Plate:	Strata DE 400 µL 96-Well Plate
Part No.:	8E-S325-5GB
Load:	Pre-treated sample onto plate (apply vacuum or positive pressure to pull/push sample into sorbent)
Wait:	5 minutes
Elute:	3x 600 µL Dichloromethane (DCM) or 3x 600 µL Ethyl Acetate
Apply:	Vacuum or apply positive pressure at 5-10" Hg for 10 seconds
Dry:	Sample under slow stream of Nitrogen at 30 °C
Reconstitute:	200 µL Acetonitrile/Water (20:80)

LC-MS/MS Conditions

Column: Kinetex[®] 2.6 µm Biphenyl
Dimensions: 50 x 2.1mm
Part No.: 00B-4622-AN
Guard Cartridge: SecurityGuard[™] ULTRA Biphenyl
Guard Part No.: AJ0-9209
Mobile Phase: A: 0.1 % Formic acid in water
 B: 0.1 % Formic acid in Acetonitrile

Gradient: Time (min)	% B
0	20
3	95
3.5	95
3.51	20
6	20

Injection Volume: 5 µL
Flow Rate: 0.5 mL/min
Detection: MS/MS (SCIEX API 4000[™]), ESI+

Recovery Values and % CVs

Elution Solvent:	Dichloromethane		Ethyl Acetate	
	%Recovery	% RSD (n=4)	%Recovery	% RSD (n=10)
β-Methasone	92	4	98	6
Cortisone	96	10	96	8
Coritcosterone	92	3	74	10
Cortisone Acetate	90	12	112	12
Triamcinolone	13	8	92	9
Prednisone	94	7	93	10
Testosterone	95	5		

*Testosterone was not extracted using Ethyl Acetate



APPLICATIONS

Ordering Information

Kinetex[®] Core-Shell LC Columns

Kinetex 2.6 µm Minibore Columns (mm)				SecurityGuard [™] ULTRA Cartridges [†]	
Phases	30 x 2.1	50 x 2.1	100 x 2.1	150 x 2.1	3/pk
Biphenyl	00A-4622-AN	00B-4622-AN	00D-4622-AN	00F-4622-AN	AJO-9209 for 2.1 mm ID

Kinetex 2.6 µm MidBore [™] Columns (mm)				SecurityGuard ULTRA Cartridges [†]	
Phases	50 x 3.0	100 x 3.0	150 x 3.0	3/pk	
Biphenyl	00B-4622-YO	00D-4622-YO	00F-4622-YO	AJO-9208 for 3.0 mm ID	

[†] SecurityGuard ULTRA Cartridges required holder, Part No.: AJO-9000.

Other particle sizes and dimensions are available. Visit www.phenomenex.com/Kinetex for a complete list of available columns.

Australia

t: +61 (0)2-9428-6444
f: +61 (0)2-9428-6445
auiinfo@phenomenex.com

Austria

t: +43 (0)1-319-1301
f: +43 (0)1-319-1300
anfrage@phenomenex.com

Belgium

t: +32 (0)2 503 4015 (French)
t: +32 (0)2 511 8666 (Dutch)
f: +31 (0)30-2383749
beinfo@phenomenex.com

Canada

t: +1 (800) 543-3681
f: +1 (310) 328-7768
info@phenomenex.com

China

t: +86 400-606-8099
f: +86 (0)22 2532-1033
phen@agela.com

Denmark

t: +45 4824 8048
f: +45 4810 6265
nordicinfo@phenomenex.com

Finland

t: +358 (0)9 4789 0063
f: +45 4810 6265
nordicinfo@phenomenex.com

France

t: +33 (0)1 30 09 21 10
f: +33 (0)1 30 09 21 11
franceinfo@phenomenex.com

Germany

t: +49 (0)6021-58830-0
f: +49 (0)6021-58830-11
anfrage@phenomenex.com

India

t: +91 (0)40-3012 2400
f: +91 (0)40-3012 2411
indiainfo@phenomenex.com

Ireland

t: +353 (0)1 247 5405
f: +44 1625-501796
eireinfo@phenomenex.com

Italy

t: +39 051 6327511
f: +39 051 6327555
italiainfo@phenomenex.com

www.phenomenex.com

Phenomenex products are available worldwide. For the distributor in your country, contact Phenomenex USA, International Department at international@phenomenex.com

Luxembourg

t: +31 (0)30-2418700
f: +31 (0)30-2383749
nlinfo@phenomenex.com

Mexico

t: 01-800-844-5226
f: 001-310-328-7768
tecnicomx@phenomenex.com

The Netherlands

t: +31 (0)30-2418700
f: +31 (0)30-2383749
nlinfo@phenomenex.com

New Zealand

t: +64 (0)9-4780951
f: +64 (0)9-4780952
nzinfo@phenomenex.com

Norway

t: +47 810 02 005
f: +45 4810 6265
nordicinfo@phenomenex.com

Puerto Rico

t: +1 (800) 541-HPLC
f: +1 (310) 328-7768
info@phenomenex.com

Spain

t: +34 91-413-8613
f: +34 91-413-2290
espinfo@phenomenex.com

Sweden

t: +46 (0)8 611 6950
f: +45 4810 6265
nordicinfo@phenomenex.com

United Kingdom

t: +44 (0)1625-501367
f: +44 (0)1625-501796
ukinfo@phenomenex.com

USA

t: +1 (310) 212-0555
f: +1 (310) 328-7768
info@phenomenex.com

All other countries Corporate Office USA

t: +1 (310) 212-0555
f: +1 (310) 328-7768
info@phenomenex.com

Strata[®] DE Supported Liquid Extraction

Part No.	Description	Unit
8E-S325-FGB	Strata DE SLE 200 µL 96-Well Plate	2/pk
8E-S325-5GB	Strata DE SLE 400 µL 96-Well Plate	2/pk
8B-S325-KDG	Strata DE SLE 12 cc Tube	20/pk
8B-S325-VFF	Strata DE SLE 60 cc Tube	16/pk

Presston[™] 100 Positive Pressure Manifold

Part No.	Description
AHO-9334	Presston 100 Positive Pressure Manifold, 96-Well Plate
AHO-9342	Presston 100 Positive Pressure Manifold, 1 mL Tube Complete Assembly
AHO-9347	Presston 100 Positive Pressure Manifold, 3 mL Tube Complete Assembly
AHO-9343	Presston 100 Positive Pressure Manifold, 6 mL Tube Complete Assembly

Presston 100 Tube Adapter Kits (for AHO-9334)

Part No.	Description
AHO-9344	1 mL Tube Adapter Kit
AHO-9345	3 mL Tube Adapter Kit
AHO-9346	6 mL Tube Adapter Kit



WARRANTY Phenomenex warrants that for a period of 12 months following delivery, the Presston 100 Positive Pressure Manifold you have purchased will perform in accordance with the published specifications and will be free from defects in materials or workmanship. In the event that the Presston 100 Positive Pressure Manifold does not meet this warranty, Phenomenex will repair or replace defective parts. Please visit www.phenomenex.com/Presston for complete warranty information.



If Phenomenex products in this technical note do not provide at least an equivalent separation as compared to other products of the same phase and dimensions, return the product with comparative data within 45 days for a FULL REFUND.

Terms and Conditions

Subject to Phenomenex Standard Terms and Conditions, which may be viewed at <http://www.phenomenex.com/TermsAndConditions>.

Trademarks

Strata and Kinetex are registered trademarks and Presston, MidBore, and SecurityGuard are trademarks of Phenomenex. API 4000 is a trademark of AB SCIEX, Pte. Ltd. AB SCIEX is being used under license.

Disclaimer

FOR RESEARCH PURPOSES ONLY. Not for use in diagnostic procedures.

© 2017 Phenomenex, Inc. All rights reserved.